



HAL
open science

On nonlinearized inversion procedures and their application to breast imaging

Yingying Qin, Marc Lambert, Thomas Rodet, Dominique Lesselier

► **To cite this version:**

Yingying Qin, Marc Lambert, Thomas Rodet, Dominique Lesselier. On nonlinearized inversion procedures and their application to breast imaging. XXXIII General Assembly and Scientific Symposium (GASS) of the International Union of Radio Science, Aug 2020, Rome, Italy. pp.B06-01. hal-02906382

HAL Id: hal-02906382

<https://centralesupelec.hal.science/hal-02906382>

Submitted on 30 Aug 2020

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.



On nonlinearized inversion procedures and their application to breast imaging

Y. Qin⁽¹⁾⁽²⁾, M. Lambert⁽³⁾, T. Rodet⁽²⁾, and D. Lesselier⁽¹⁾

(1) Université Paris-Saclay, CNRS, CentraleSupélec, Laboratoire des signaux et systèmes, 91190, Gif-sur-Yvette, France

(2) Université Paris-Saclay, ENS Paris-Saclay, CNRS, Systèmes et Applications des Technologies de l'Information et de l'Énergie, 94235, Cachan, France

(3) Université Paris-Saclay, CentraleSupélec, CNRS, Laboratoire de Génie Électrique et Électronique de Paris, 91192, Gif-sur-Yvette, France

Breast imaging using electromagnetic (EM) waves is challenging, as illustrated by an abundant literature (not further discussed for brevity), yet lacks application in the field beyond prototype stage, e.g., [1]. In the present contribution one aims at the case of a pending breast (in opposition with the compressed breast of X-ray mammography) immersed in a coupling medium and probed by a series of antennas (emitters and transmitters as well) set in this medium, frequencies of their operation being chosen around a couple of GHz to achieve both enough penetration and still meaningful resolution. The spatial distribution of the EM parameters (permittivity and conductivity, accounting for the fact that frequency-dispersive tissues usually follow Debye-type laws) is to be retrieved so as potential anomalies be exhibited. A broad set of inversion procedures could be applied to tackle such an inverse scattering problem, refer, e.g., to [2], though it is obvious from earlier investigations that priors as those yielded by ultrasonic (US) probing [3] can be very useful. When the contrasts are as weak as those in US breast imaging (with speed of sound varying by no more than a few percents), the inversion indeed can be cast as a linear problem, and in practice simple reflectometric analyses led in the low MHz range (associated with mm local wavelengths) can provide boundaries (possibly blurred) between zones in the breast, even if not attuning for slowly changing parameters within a given zone. This is in contrast with the highly nonlinear aspect of the EM case, since contrasts can be high, in addition to the heterogeneity of the breast structure, plus the impedance jump at the skin.

Several solution methods are to be compared in this contribution (with aforementioned US priors) on a 4-zone 2-D model of breast (skin, fatty, fibroglandular and tumor tissues) and a MRI-derived one from the UWCEM Breast Phantom Repository, with due application of a Method of Moments (FFT being implemented in both forward and inverse problems to accelerate the computations). In a frequency-hopping setting, using the US boundary priors, the rather traditional Distorted Born Iterative Method (DBIM) and Contrast Source Inversion (CSI) method are employed (with proper tailoring), and their pros and cons compared in systematic fashion with those of a novel multi-binary contrast-source-based extension of a binary-constrained solution first introduced in [3] and of a Twofold Subspace-based Optimization Method (TSOM) as recently proposed with emphasis on imaging obstacles in a known inhomogeneous medium. Time permitting, from all such results, the potential of data fusion using both US and EM modalities in the course of the inversion itself instead of just introducing smart US priors will be considered, including the fact that one is faced with cm local wavelengths with EM and millimetric ones with US, which complexifies data fusion (US mapping of boundaries first is in comparison more straightforward).

References

- [1] M. Asefi, A. Baran, and J. LoVetri, "An experimental phantom study for air-based quasi-resonant microwave breast imaging," *IEEE Trans. Microw. Theory Tech.*, Vol. 67, no. 9, pp. 3946-3954, 2019.
- [2] X. Chen, *Computational Methods for Electromagnetic Inverse Scattering*, 1st Edition, Wiley-IEEE Press, Singapore, 2018.
- [3] N. Abdollahi, D. Kurrant, P. Mojabi, M. Omer, E. Fear, and J. LoVetri, "Incorporation of ultrasonic prior information for improving quantitative microwave imaging of breast," *IEEE J. Multiscale Multiphys. Comput. Tech.*, Vol. 4, pp. 98-110, 2019.
- [4] K. Xu, Y. Zhong, X. Chen, and D. Lesselier, "A fast integral equation-based method for solving electromagnetic inverse scattering problems with inhomogeneous background," *IEEE Trans. Antennas Propagat.*, Vol. 66, no. 8, pp. 4228-4238, 2018.
- [5] L. Souriau, B. Duchêne, D. Lesselier, and R. E. Kleinman, "Modified gradient approach to inverse scattering for binary objects in stratified media," *Inverse Probl.*, Vol. 12, pp. 463-481, 1996.